

Certificate



Certificate no.

TA 50388011 01

License Holder:

Zippy Technology Corp.
10F., No. 50, Minquan Rd.,
Shin Tien District, New Taipei City
231
Taiwan

Manufacturing Plant:

ZIPPY TECHNOLOGY CORP
4F., No. 48, Wucyuan Rd.
Wugu Dist., New Taipei City 24886
Taiwan

Test report no.: ZTW1-SCL 50098166 001

Client Reference: SPC-MHG2 series

Tested to: IEC 60601-1:2005+A1
EN 60601-1:2006+A11+A1+A12

Certified Product: (Switching Power Supply)

License Fee - Units

Type Designation	: MHG2-6400P (EMACS)	10
Rated Voltage	: AC 100-240V, 47-63Hz	
Rated Current	: 8-4A	
Output Voltage	: see constructional dataform	
Output Current	: see constructional dataform	
max. Power Output	: see constructional dataform	
max. Ambient Temperature	: 50°C	
Protection Class	: I	

This certificate is valid for maximum 5 years.

Appendix: 1

10

Licensed Test mark:



TÜV Rheinland Taiwan Ltd

Signature

Dipl.-Ing. W. Hsu



Date of Issue
(day/mo/yr)

05/10/2017

Certificate



Certificate no.

TA 50388011 02

License Holder:

Zippy Technology Corp.
10F., No. 50, Minguang Rd.,
Shin Tien District, New Taipei City
231
Taiwan

Manufacturing Plant:

ZIPPY TECHNOLOGY CORP
4F., No. 48, Wucyuan Rd.
Wugu Dist., New Taipei City 24886
Taiwan

Test report no.: ZTW1-SCL 50098166 001

Client Reference: SPC-MHG2 series

Tested to:

IEC 60601-1:2005+A1
EN 60601-1:2006+A11+A1+A12

Certified Product: (Switching Power Supply)

License Fee - Units

as page 01

Addition

Type Designation	: MHG2-6350P (EMACS)	1
	MHG2-6300P (EMACS)	1
Rated Voltage	: AC 100-240V, 47-63Hz	
Rated Current	: 7-3.5A	

This certificate is valid for maximum 5 years.

Appendix: 1

2

Licensed Test mark:



TÜV Rheinland Taiwan Ltd

Signature

Dipl.-Ing. W. Hsu



Date of Issue
(day/mo/yr)

05/10/2017