



Ref. Certif. No.

DE 2-022465

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT
(IECEE) CB SCHEME

CB TEST CERTIFICATE

Product

Switching Power Supply

Name and address of the applicant

Zippy Technology Corp.
10F, No. 50, Minquan Rd.
Shin Tien District, New Taipei City, Taiwan 231

Name and address of the manufacturer

Zippy Technology Corp.
10F, No. 50, Minquan Rd.
Shin Tien District, New Taipei City, Taiwan 231

Name and address of the factory

ZIPPY TECHNOLOGY CORP
4F, No. 48 Wucyuan Rd.
Wugu Dist., New Taipei City 24886, Taiwan

Ratings and principal characteristics

AC 100-240V; 47-63Hz; 6-3A for MHU2-5400V;
8-4A for MPSM-5500V; 10-5A for MPSM-5600V; Class I
for output ratings refer to test report page 6

Trademark (if any)

EMACS

Customer's Testing Facility (CTF) Stage used

N/A

Model / Type Ref.

MHU2-5400V; MPSM-5600V; MPSM-5500V

Additional information (if necessary may
also be reported on page 2)

Models differ in output current, air-flow of fan, direction
of air-flow, configuration of metal chassis, heatsink of
BD1, ratings of primary and secondary components; for
details refer to test report

A sample of the product was tested and
found to be in conformity with

IEC 60601-1:2005 + A1
for national deviations see test report

As shown in the Test Report Ref. No. which
forms part of this Certificate

50094335 001

This CB Test Certificate is issued by the National Certification Body



TÜV Rheinland LGA Products GmbH
Tillystraße 2 · 90431 Nürnberg, Germany
Phone + 49 221 806-1371
Fax + 49 221 806-3935
Mail: cert-validity@de.tuv.com
Web: www.tuv.com



Date: 27.10.2017

Signature:

Dipl.-Ing. W. Hsu