

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT
(IECEE) CB SCHEME

CB TEST CERTIFICATE

Product

Medical Power Supply

Name and address of the applicant

Zippy Technology Corp
10F., No. 50, Minquan Rd,
Xindian District, New Taipei 231, Taiwan

Name and address of the manufacturer

Zippy Technology Corp
10F., No. 50, Minquan Rd,
Xindian District, New Taipei 231, Taiwan

Name and address of the factory

ZIPPY TECHNOLOGY CORP
4F., No. 48 Wucyuan Rd.
Wugu Dist., New Taipei City 24886, Taiwan

Ratings and principal characteristics

AC 100-240V; 47-63Hz; 5-2.5A (xxx = 300); 6-3A (xxx = 400);
8-4A (xxx = 500); Class I;
for output ratings refer to test report page 6

Trademark (if any)

EMACS

Customer's Testing Facility (CTF) Stage used

N/A

Model / Type Ref.

MHG2-5xxxG
(xxx = 300, 400, 500 to denote different input current
and output ratings)Additional information (if necessary may
also be reported on page 2)Models differ in model name, input and output ratings,
sources of main transformer T5, the ratings of components;
for details refer to the test reportA sample of the product was tested and
found to be in conformity withIEC 60601-1:2005+A1
for national deviations see test reportAs shown in the Test Report Ref. No. which
forms part of this Certificate

50199564 001

This CB Test Certificate is issued by the National Certification Body



TÜVRheinland®

TÜV Rheinland LGA Products GmbH
Tillystraße 2 · 90431 Nürnberg, Germany
Phone + 49 221 806-1371
Fax + 49 221 806-3935
Mail: cert-validity@de.tuv.com
Web: www.tuv.com

Date: 08.03.2019

Signature:

Dipl.-Ing. W. Hsu

