

Certificate



Certificate no.

TA 50430343 01

License Holder:

Zippy Technology Corp
10F., No. 50, Minquan Rd,
Xindian District, New Taipei 231
Taiwan

Manufacturing Plant:

ZIPPY TECHNOLOGY CORP
4F., No. 48, Wucyuan Rd.
Wugu Dist., New Taipei City 24886
Taiwan

Test report no.: ZTW1- 50199571 001

Client Reference: 107-0917/PSE

Tested to: IEC 60601-1:2005+A1
EN 60601-1:2006+A11+A1+A12

Certified Product: (Medical Power Supply)

License Fee - Units

Type Designation	: MHG2-5xxxG (EMACS)	10
xxx stands for	: 300, 400, 500	1
Rated Voltage	: AC 100-240V, 47-63Hz	
Rated Current	: see constructional dataform	
Output Voltage	: see constructional dataform	
Output Current	: see constructional dataform	
max. Power Output	: see constructional dataform	
max. Ambient Temperature:	1) +50°C for cooling fan outward 2) +40°C for cooling fan inward	
Protection Class	: I	

This certificate is valid for maximum 5 years.

Appendix: 1

11

Licensed Test mark:



IEC 60601-1
EN 60601-1

www.tuv.com
ID 1419066648

TÜV Rheinland Taiwan Ltd
Signature

Dipl.-Ing. W. Hsu

Date of Issue
(day/mo/yr)
08/03/2019