



Ref. Certif. No.

DE 2-023803

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT  
(IECEE) CB SCHEME

CB TEST CERTIFICATE

Product

Switching Power Supply

Name and address of the applicant

Zippy Technology Corp  
10F., No. 50, Minquan Rd,  
Xindian District, New Taipei 231, Taiwan

Name and address of the manufacturer

Zippy Technology Corp  
10F., No. 50, Minquan Rd,  
Xindian District, New Taipei 231, Taiwan

Name and address of the factory

ZIPPY TECHNOLOGY CORP  
4F., No. 48 Wucyuan Rd.  
Wugu Dist., New Taipei City 24886, Taiwan

Ratings and principal characteristics

AC 100-240V; 47-63Hz; 3.5-2A or 4-2A or 4.5-2A or 6-3A;  
Class I  
for output ratings refer to test report no. 50094337 002  
page 5

Trademark (if any)

EMACS

Customer's Testing Facility (CTF) Stage used

N/A

Model / Type Ref.

MP1S-xV; AD-60040-SHM  
(x = 5220, 5250, 5300, 6300, 6400 to denote different  
input current and output)

Additional information (if necessary may  
also be reported on page 2)

-update sources of triple insulated wire of transformer  
(T1, T2); for details see test report;  
-see also test report ref. no.50094337 001;  
-supersede previously issued certificate DE 2-022322

A sample of the product was tested and  
found to be in conformity with

IEC 60601-1:2005+A1  
for national deviations see test report

As shown in the Test Report Ref. No. which  
forms part of this Certificate

50094337 002

This CB Test Certificate is issued by the National Certification Body



TÜV Rheinland LGA Products GmbH  
Tillystraße 2 · 90431 Nürnberg, Germany  
Phone + 49 221 806-1371  
Fax + 49 221 806-3935  
Mail: cert-validity@de.tuv.com  
Web: www.tuv.com



Date: 05.09.2018

Signature:

Dipl.-Ing. W. Hsu