

Certificate



Certificate no.

TA 50484548 01

License Holder:

Zippy Technology Corp
10F., No. 50, Minquan Rd,
Xindian District, New Taipei 231
Taiwan

Manufacturing Plant:

Refer to latest revision
of the annex list of factories

Test report no.: ZTW1- 60417692 001

Client Reference: SC-08072020-CB+TUV

Tested to: IEC 60601-1:2005+A1
EN 60601-1:2006+A11+A1+A12

Certified Product: (Medical Power Supply)

License Fee - Units

Type Designation	: MP1P-5300G (EMACS)	10
Rated Voltage	: AC 110-240V, 47-63Hz	
Rated Current	: 4-2.5A	
Output Voltage	: see constructional dataform	
Output Current	: see constructional dataform	
max. Power Output	: see constructional dataform	
max. Ambient Temperature	: 50°C	
Protection Class	: I	

This certificate is valid for maximum 5 years.

Appendix: 1

10

Licensed Test mark:



IEC 60601-1
EN 60601-1

www.tuv.com
ID 1111231269

TÜV Rheinland Taiwan Ltd

Signature



Dipl.-Ing. (FH) A. Klinker

Date of Issue

(day/mo/yr)

16/11/2020

Anlage Fertigungsstättenliste
/Attachment List of Factories



TA 50484548 0001

1 ZIPPY TECHNOLOGY CORP
4F., No. 48, Wucyuan Rd.
Wugu Dist., New Taipei City 24886
Taiwan



Dieser Anhang ersetzt den Vorgänger vom/
This annex replaces the previous annex dated

Datum / Date 16.11.2020



Zertifizierungsstelle
/ Certification Body

A handwritten signature in blue ink, appearing to be "A. Klinker", is written over a horizontal line.

Dipl.-Ing. (FH) A. Klinker